



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Toll Company

3005 Niagara Lane North, Plymouth, MN 55447

*(Hereinafter called the Organization) and hereby declares that Organization is accredited
in accordance with the recognized International Standard:*

ISO/IEC 17025:2005

This accreditation demonstrates technical competence for a defined scope and the
operation of a laboratory quality management system
(as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

Chemical Testing ***(As detailed in the supplement)***

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.


For PJLA:

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

<i>Initial Accreditation Date:</i>	<i>Issue Date:</i>	<i>Accreditation No.:</i>	<i>Certificate No.:</i>
April 9, 2009	May 13, 2013	64630	L13-100

*The validity of this certificate is maintained through ongoing assessments based
on a continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjilabs.com*



Certificate of Accreditation: Supplement

Toll Company

3005 Niagara Lane North, Plymouth, MN 55447
Chuck Allard Phone: 763-551-5300

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Chemical	High Pressure and Cryogenic Gases	Trace Moisture Concentration	Electrolytic Hygrometer	0.000 01 % mol fraction 0.01 % to mol fraction (0.000 01 % mol fraction LoD)
		Trace Hydrocarbon Concentration	Flame Ionization Detector	0.000 01 % mol fraction to 10 % mol fraction (0.000 001 % mol fraction LoD)
		Gas Mixture Concentration	Electrochemical Oxygen Analyzer	0.000 01 % mol fraction to 23.0 % mol fraction (0.000 001 % mol fraction LoD)
		Trace Gas Mixture Concentration	Gas Chromatograph with Pulsed Discharge-Ionization Detector-PDID	0.000 01 % mol fraction to 0.1 % mol fraction (0.000 01 % mol fraction LoD)
		Percent Oxygen Concentration	Paramagnetic Oxygen Analyzer	1.0 % mol fraction to 100 % mol fraction (0.1 % mol fraction LoD)
		Binary Gas Mixture Concentration	Thermal Conductivity Detector	1.0 % mol fraction to 100 % mol fraction (0.1 % mol fraction LoD)
		Gas Mixture Concentration	Gravimetric Balance	0.000 01 % mol fraction to 100 % mol fraction (0.000 01 % mol fraction LoD)



Perry Johnson Laboratory Accreditation, Inc.



May 13, 2013

Chuck Allard
Toll Company
3005 Niagara Lane North
Plymouth, MN 55447

Dear Mr. Allard:

This letter is to confirm that you have successfully completed your reaccreditation assessment. A certificate has now been granted and posted on our website. As you are aware, PJLA will no longer be issuing expiration dates on our certificates. Your certificate # **L13-100** will remain valid as long as you continue to maintain your annual assessments and reaccreditation assessments as stated in your customer agreement with PJLA. At this time, we have confirmed that your annual assessments will be conducted during the month of **March** each calendar year. This will include an interim surveillance assessment and a full system reassessment to be completed by **March 2015**. Once your reassessment is conducted and approved by our accreditation committee a revised status letter will be provided to you. Please allow PJLA at least 120 days from your assessment due date to issue this letter.

Please feel free to release this letter to any interested parties as confirmation of your certificate validity. Also, please remind them that your certificate is posted on our website at all times. Any changes in regards to your accreditation status will be reflected on our website.

We would like to thank you for your patronage over the past years and look forward to continuously serving your accreditation needs in the future. If we can assist you any further, please feel free to contact us at any time.

Sincerely,

Tracy Szerszen
President/Operations Manager